



SI – 3000 CELJE – Opekarniška cesta 26  
Tel.: +386 (0) 3 425-62-00 Fax: +386 (0) 3 425-62-10  
E-mail: [info@interdent.cc](mailto:info@interdent.cc)  
[http:// www.interdent.cc](http://www.interdent.cc)

## IZJAVA O SKLADNOSTI DECLARATION OF CONFORMITY

Podjetje/Company: **INTERDENT® d.o.o.**  
Naslov/Address: **Opekarniška cesta 26, SI - 3000 CELJE**

Na lastno odgovornost izjavljamo, da sledeči proizvodi, razreda IIa (pravilo 5)  
*We herewith declare on our sole responsibility that the following Class IIa Products (rule 5)*

<b>GENERICNO IME / GENERIC NAME</b>	<b>CAD/CAM DISKI PMMA / CAD/CAM DISCS PMMA</b>
<b>TRGOVSKO IME / TRADE NAME</b>	CC DISK PMMA (predbarvani / <i>preshade</i> , 3-slojni / <i>3- layers</i> , Multilayer)
<b>UMDNS / GMDN</b>	<b>45236</b>

ustrezajo bistvenim zahtevam Direktive o medicinskih pripomočkih 93/42/EGS.  
comply with essential requirements of the Medical Devices Directive 93/42 EEC.

Postopek ugotavljanja skladnosti: Dodatek II (brez točke 4) Direktive o medicinskih pripomočkih 93/42/EGS, datum izdaje: 19. 05. 2021, številka registracije: HD 1076832-1, veljavnost certifikata: 26. 05.2024

*Conformity assessment procedure: Annex II (without point 4) of Medical Device Directive 93/42/EEC, date of issue: 19<sup>th</sup> May, 2021, registration No: HD 1076832-1, certificate validity: 26<sup>th</sup> May 2024*

Priglašeni organ za ugotavljanje skladnosti / *Notified body:*

TÜV Rheinland LGA Products GmbH, Tillystrasse 2, D – 90431 Nürnberg – številka / *number* **0197**

### HARMONIZIRANI IN OSTALI STANDARDI / *HARMONISED AND OTHER STANDARDS:*

EN ISO 13485:2016 Medicinski pripomočki – Sistem vodenja kakovosti – Zahteve za zakonodajne namene / *Medical devices – Quality management systems – Requirements for regulatory purpose*

EN ISO 14971:2012 Medicinski pripomočki-Uporaba obvladovanja tveganja pri medicinskih pripomočkih / *Medical devices - Application of risk management to medical devices*

EN ISO 10993-1:2020 Biološko vrednotenje medicinskih pripomočkov – 1. del: Ocena in preskusi znotraj ocene tveganja / *Biological Evaluation of Medical Devices- Part 1: Evaluation and testing within a risk management process*

EN ISO 7405:2018 Zobozdravstvo – Ocena biokompatibilnosti medicinskih pripomočkov v zobozdravstvu / *Dentistry – Evaluation of biocompatibility of medical devices used in dentistry*

EN ISO 10993-5:2009: Biološko vrednotenje medicinskih pripomočkov – 5 del: Preskusi za ugotavljanje citotoksičnosti in vitro / *Biological Evaluation of Medical Devices- Part 5: 2009 Testing of cytotoxicity: In vitro methods*

EN ISO 10993-10:2013 Biološko vrednotenje medicinskih pripomočkov – 10.del: Preskus draženja in preobčutljivosti kože / *Biological evaluation of medical devices – Part 10:Test for irritation and skin sensitization.*

EN 62366-1:2015 Medicinski pripomočki – Del 1: Uporaba inženiringa uporabnosti medicinskih pripomočkov/ *Medical devices – Application of usability engineering to medical devices*

EN ISO 15223-1:2016: Medicinski pripomočki – Simboli za označevanje medicinskih pripomočkov, označevanje in podatki, ki jih mora podati dobavitelj- 1.del: Splošne zahteve / *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – part 1: general requirements*

EN 1641:2009 Zobozdravstvo - Medicinski pripomočki za zobozdravstvo - Materiali / *Dentistry – Part 1: Medical devices for dentistry - Materials.*

EN 1041:2008+A1:2013 Informacije proizvajalca za medicinske pripomočke / *Information supplied by the manufacturer of medical devices*

EN ISO 10477:2020 Zobozdravstvo – Polimerni materiali za krone in mostičke / *Dentistry – Polymer based crown and bridge materials*

EN ISO 20795-1:2013 Zobozdravstvo – Osnovni polimeri – 1. Del: Osnovni polimeri za proteze / *Dentistry- Base polymers – Part 1: denture base polymers*

ISO 178:2010 Plastika – Določanje lastnosti plastičnosti / *Plastic – Determination of plastic properties*

EN ISO 9001:2015- Sistem vodenja kakovosti – zahteve / *Quality management system – Requirements*

Veljavnost izjave o skladnosti je vezana na spremembo medicinskega pripomočka ali na veljavnost certifikata priglašene organa. / *The validity of declaration of conformity is linked to a change in medical device or on validity of certificate issued by notified body.*

Celje, 25.08.2021

Anja Mavrič, univ.dipl.biol.



**Place, Date**

**Responsible person for MD and technical files**

**Signature:**

Verzija / Version: 2